

AUG 23 2010

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130305</u>	
Name <u>Anthony Wallace</u>		ID# <u>6555</u>	Date <u>8/10/10</u>
A	Agency <u>Hoonah Police Dept.</u>		Phone # <u>945-3655</u>
Instrument Location <u>Hoonah Police Dept</u>			
Alco S/N <u>4301532</u>		Target Value <u>.082</u>	High Pressure <u>500</u>
B	Alco Test Value Average <u>.088</u>		<u>.088</u>
	1 <sup>st</sup> Alco	2 <sup>nd</sup> Alco	
Signature <u>[Signature]</u>			
(OVER)			

BMB  
8/27/10


I, Nita J. Bolz, after being first duly sworn, depose and state as follows: (Do Not write in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 23<sup>RD</sup> day of SEPT, 2010.

[Signature] (Notary Seal Stamp)  
Daniel L. Nicholas  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130305

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130305

AUGUST 10, 2010

OPERATOR'S NAME:  
WALLACE ANTHONY  
OPERATOR'S NUMBER: 6555  
SUBJECT'S LAST NAME:

UOC  
SUBJECT'S FIRST NAME/MI :  
UOC

O.L. #: UOC  
DEPT/AGENCY: HHHL  
CASE/REPORT: UOC  
TEST TYPE: U  
ALCO TARGET VALUE: .002  
ALCO S/N: X301532

### — BREATH ANALYSIS —

.002 ADJUSTED FOR 30.19 in  
ALCO TARGET .002 22:02  
BLANK TEST .000 22:03  
INTERNAL STANDARD .000 22:03  
ALCO TV 30.19 in .000 22:04  
BLANK TEST .000 22:04  
SUBJECT SAMPLE .000 22:05  
BLANK TEST .000 22:06  
ALCO TV 30.19 in .000 22:06  
BLANK TEST .000 22:07

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130305

AUGUST 10, 2010  
TIME 22:00

### — DIAGNOSTIC CHECK —

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 01/08/89  
HEATERS  
SAMPLE CHAMBER: 40c  
BREATH TUBE: 40c  
BAROMETER: 30.19 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

### PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGHI  
HIJKLMNOPQRSTUVWXYZ[\]^\_`abcdefghijklmnopqrstuvwxyz{|}~